



Aug 26th, 2024
Cosel Co., Ltd.
WS Design Section

EMI/EMS Test Result
According to IEC60601-1-2 4th Edition (EMS)

Model Name : UMCS30F series

Approved : Takashi Kajii

The EUT is operated with following condition during EMI/EMS test.

Input Voltage : 230,240VAC / 50Hz
Output Current : Rated Current
Ambient Temperature : 25°C ± 10°C

Prepared : Yuya Sakai

| # | Subject | Reference standard | Test Condition | Criteria *1 | Result |
|----|---------|--|---|-------------|-------------------------|
| 1 | EMI | Conducted Emission | EN55011, EN55032 Class B CISPR11, CISPR32 Class B FCC Part15, FCC Part18 Class B | - | Pass |
| 2 | | Radiated Emission | EN55011, EN55032 Class B CISPR11, CISPR32 Class B FCC Part15, FCC Part18 Class B | - | Pass |
| 3 | | Harmonic Current | IEC61000-3-2 Class A | - | Pass |
| 4 | EMS | Electrostatic discharge immunity test | IEC61000-4-2 Contact Discharge : Level 4 (8kV) Air Discharge : Level 4 (15kV) | A | Pass |
| 5 | | Radiated, radio-frequency, electromagnetic field immunity test | IEC61000-4-3 10V/m : (80MHz~2.7GHz) 80% Amplitude modulated | A | Pass |
| 6 | | Electrical fast transient / Burst immunity test | IEC61000-4-4 Level 4 (4kV) Repetition Rate : 5kHz and 100kHz | A | Pass |
| 7 | | Surge immunity test | IEC61000-4-5 Line to Line : Level 4 (2kV) | A | Pass |
| 8 | | Immunity to conducted disturbances, induced by radio-frequency fields | IEC61000-4-6 Voltage Level (e.m.f.) : Level 3 (10Vrms) | A | Pass |
| 9 | | Power frequency magnetic field Immunity test | IEC61000-4-8 Magnetic Field Strength : Level 4 (30A/m) | A | Pass |
| 10 | | Voltage dips, short interruptions and voltage variations immunity test | IEC61000-4-11 (1) 100% dip for 10ms, 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315° (2) 100% dip for 20ms, 0° (3) 30% dip for 500ms, 0° | A A A | Pass *2 Pass Pass |
| | | | (4) 100% dip for 5 seconds (short interruption) | B | Pass |

*1 Definition of Criteria
Criteria A : (1) No output voltage drop with control circuit failure.
(2) No protection circuit and other circuit malfunction.

Criteria B : (1) The output voltage is temporary degradation of performance.
It recovers its normal performance without operator intervention.
(2) No protection circuit and other circuit failure.

*2 Output Current:70% or less of rated current(at 100VAC)

<Notes>
Power supply shall not determine the final equipment performance against EMS test. Therefore we confirmed the output voltage performance only. EMS test should be performed as a final product.